



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
[www.uspto.gov](http://www.uspto.gov)

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/730,276	12/09/2003	Alain Tormier	14542	2602
293	7590	10/18/2005	EXAMINER	
Ralph A. Dowell of DOWELL & DOWELL P.C. 2111 Eisenhower Ave. Suite 406 Alexandria, VA 22314			MCKANE, ELIZABETH L	
		ART UNIT		PAPER NUMBER
		1744		

DATE MAILED: 10/18/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

<b>Office Action Summary</b>	<b>Application No.</b>	<b>Applicant(s)</b>	
	10/730,276	TORNIER, ALAIN	
	Examiner Leigh McKane	Art Unit 1744	

*-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --*  
**Period for Reply**

**A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.**

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) Responsive to communication(s) filed on \_\_\_\_.
- 2a) This action is FINAL.                    2b) This action is non-final.
- 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) Claim(s) 1-10 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_ is/are withdrawn from consideration.
- 5) Claim(s) \_\_\_\_ is/are allowed.
- 6) Claim(s) 1-10 is/are rejected.
- 7) Claim(s) \_\_\_\_ is/are objected to.
- 8) Claim(s) \_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) The specification is objected to by the Examiner.
- 10) The drawing(s) filed on 09 December 2003 is/are: a) accepted or b) objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) All    b) Some \* c) None of:
  1. Certified copies of the priority documents have been received.
  2. Certified copies of the priority documents have been received in Application No. \_\_\_\_.
  3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- |   |   |
|---|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)  | 4) <input type="checkbox"/> Interview Summary (PTO-413)                     |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)  | Paper No(s)/Mail Date. ____ .   |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)<br>Paper No(s)/Mail Date <u>120903</u> . | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
|   | 6) <input type="checkbox"/> Other: ____ .                                   |

***Claim Rejections - 35 USC § 103***

1. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

2. Claims 1-10 are rejected under 35 U.S.C. 103(a) as being unpatentable over Nicolais (EP 982236) in view of Hamilton et al (EP 737481) in view of Ahlqvist et al (US 5881534).

With respect to claims 1-3 and 8-10, Nicolais teaches a process for the sterile packaging of a prosthetic implant **10** wherein the implant is placed in a flexible, gas-impermeable sachet **18** under vacuum and the sachet **18** is heat sealed. The sachet containing the implant is then placed within a flexible gas-impermeable envelope **20** which is also heat sealed. This sealed envelope **20** containing the sachet and implant is folded upon itself and placed within a rigid outer container **28** to protect the implant. See Figure 2; paragraphs [0012]-[0027]. Nicolais is silent with respect to the implant (hip joint prosthesis) being made of polyethylene or to forming an inert gaseous atmosphere within the envelope before sealing.

Hamilton et al discloses that it was known in the art at the time of the invention to fabricate artificial joints from polymeric materials, such as ultrahigh molecular weight polyethylene and to sterilize these joints using radiation. See Abstract; page 1, lines 7-10. It would have been obvious to one of ordinary skill in the art to employ the sterile packaging method of Nicolais to package and sterilize implants fabricated from polyethylene since they are

Art Unit: 1744

safely sterilized by radiation and since polyethylene is a common material from which artificial joints are fabricated.

Ahlqvist et al teaches that when either an article or the packaging in which the article is enclosed is fabricated from a polymer, such as polyethylene, it is necessary to remove oxygen from the atmosphere surrounding the polyethylene so that during radiation the formation of free radicals is minimized. To remove the oxygen, Ahlqvist et al discloses that the article or container be surrounded by an inert gas (nitrogen). See col.5, line 60 to col.6, line 52.

Since Nicolais alone teaches forming the envelope **20** of a polymeric material, such as polyethylene, it would have been obvious to form an inert gas atmosphere within the envelope **20** of Nicolais before sealing, thus minimizing free radical damage to the envelope.

As to claim 4, Nicolais teaches that the sachet **18** may be formed of laminates of different materials (paragraphs [0020]-[0021]) but does not teach a laminate containing aluminum for the sachet **18**. Ahlqvist et al discloses a gas-impermeable container suitable for radiation sterilization that preferably contains an aluminum layer. See col.6, lines 41-47. It would have been obvious to one of ordinary skill in the art to choose a gas-impermeable packaging material known in the art to be sealable, stable over long periods of storage, and capable of withstanding irradiation. As the packaging material of Ahlqvist et al meets these requirements and as Nicolais is not limited to a particular packaging material, it would have been obvious to one of ordinary skill in the art to choose the aluminum laminate packaging material of Ahlqvist et al for the sachet **18** of Nicolais.

With respect to claim 5, Nicolais teaches that the envelope **20** can be fabricated from "flexible polymeric films" that are gas impermeable. Suggested materials include polyethylene

Art Unit: 1744

and nylon (polyamide). See paragraph [0023]. However, Nicolais does not disclose a film containing both nylon and polyethylene. Hamilton et al teaches sealed, gas-impermeable packaging material that is irradiated for sterilization of the articles within. Suggested materials includes a multilayered film containing both nylon and polyethylene. See page 3, lines 6-14. As this packaging material fulfills the requirements of Nicolais (gas-impermeable, sealable, and radiation sterilizable), one would have found it obvious to use the packaging material of Hamilton et al. for the envelope **20** of Nicolais.

As to claims 6 and 7, the combination of Nicolais with Ahlqvist et al teaches removing oxygen from the envelope **20** (presumably by vacuum) and filling the envelope with an inert gas. Ahlqvist et al does not disclose a pressure to which one should fill the inert gas. However, as Nicolais teaches keeping the sachet **18** under a vacuum, it would have been obvious to also maintain the inert gas-filled envelope **20** below atmospheric. Moreover, in order to control the introduction of inert gas into the envelope it is deemed obvious to employ suitable control means, such as calibration, for accuracy.

### *Conclusion*

3. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Leigh McKane whose telephone number is 571-272-1275. The examiner can normally be reached on Monday-Thursday (5:30 am-2:00 pm).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, John Kim can be reached on 571-272-1142. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 1744

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

*Leigh McKane*  
**Leigh McKane**  
**Primary Examiner**  
**Art Unit 1744**

elm  
16 October 2005